



CERTIFICATE OF ANALYSIS
CITRANOX
FOR BECKMAN COULTER (BCI)

Date Of Manufacture: 06/9/2016
Manufactured by: Alconox, Inc.
Manufacturer Lot number: C9F6
BCI Lot number: 622832
Manuf. Cat. No.: 1801B
BCI Part No: 81912
Shelf Life Expiration Date: 05/31/2018 To meet BC requirements
Storage Conditions: Closed upright and in a cool dry place (15 - 30 deg C or 60 - 90 deg F).

<u>Value</u>	-	<u>Result</u>	<u>Acceptance Range</u>
Appearance:		OK	Clear pale olive colored liquid
pH: (1% solution in DI water)		2.7	2.0-3.0
Surface Tension: (1% solution Dyne/cm)		31.1	29.0 – 33.0
Specific Gravity		1.118	1.090 – 1.140
Number of Bar Code Labels Used for Labels		720	Not Applicable
Number of Bar Code Labels Used for Records		3	Not Applicable
Number of Bar Code Labels Discarded		26	Not Applicable

Malcolm McLaughlin, QA

REF 81912

2018-05-31

LOT 622832



+H628819121A



+\$\$2053118622832A0



(01)15099590223571
(17)180531
(10)622832

Alconox, Inc.

The Leader in Critical Cleaning

30 Glenn St. White Plains NY 10603 USA

Tel 914.948.4040 • Fax 914-948-4088

www.alconox.com • cleaning@alconox.com