



**CERTIFICATE OF ANALYSIS**  
**CITRANOX**  
**FOR BECKMAN COULTER (BCI)**

Date Of Manufacture: 10/2/2017  
Manufactured by: Alconox, Inc.  
Manufacturer Lot number: C9K7  
BCI Lot number: C9G7  
Manuf. Cat. No.: 1801B  
BCI Part No: 81912  
Shelf Life Expiration Date: 09/30/19 to meet BC requirements  
Storage Conditions: Closed upright and in a cool dry place (15 - 30 deg C or 60 - 90 deg F).

Value	Result	Acceptance Range
Appearance:	OK	Clear pale olive colored liquid
pH: (1% solution in DI water)	2.9	2.0-3.0
Surface Tension: (1% solution Dyne/cm)	31.5	29.0 - 33.0
Specific Gravity	1.112	1.090 - 1.140
Number of Bar Code Labels Used for Labels	1,096	Not Applicable
Number of Bar Code Labels Used for Records	6	Not Applicable
Number of Bar Code Labels Destroyed	401	Not Applicable

Malcolm McLaughlin, QA

**REF** 81912

**EXP** 2019-09-30

**LOT** C9G7



\*4628819121A\*



\*552093019C9G7AP\*



(01)15099590223571  
(17)190930  
(10)C9G7

Alconox, Inc.  
The Leader in Critical Cleaning  
30 Glenn St. White Plains NY 10603 USA  
Tel 914.948.4040 • Fax 914-948-4088  
www.alconox.com • cleaning@alconox.com