



**CERTIFICATE OF ANALYSIS**  
**CITRANOX**  
**FOR BECKMAN COULTER (BCI)**

Date Of Manufacture: 12/4/2017  
Manufactured by: Alconox, Inc.  
Manufacturer Lot number: C1M7  
BCI Lot number: C9J7  
Manuf. Cat. No.: 1801B  
BCI Part No: 81912  
Shelf Life Expiration Date: 10/31/19 to meet BC requirements  
Storage Conditions: Closed upright and in a cool dry place (15 - 30 deg C or 60 - 90 deg F).

<u>Value</u>	<u>Result</u>	<u>Acceptance Range</u>
Appearance:	OK	Clear pale olive colored liquid
pH: (1% solution in DI water)	2.8	2.0-3.0
Surface Tension: (1% solution Dyne/cm)	31.3	29.0 - 33.0
Specific Gravity	1.116	1.090 - 1.140
Number of Bar Code Labels Used for Labels	720	Not Applicable
Number of Bar Code Labels Used for Records	2	Not Applicable
Number of Bar Code Labels Destroyed	511	Not Applicable

Malcolm McLaughlin, QA

**REF** 81912

2019-10-31

**LOT** C9J7



••H628819121A•



••552103119C9J7AL•



(01)15099590223571  
(17)191031  
(10)C9J7

Alconox, Inc.  
The Leader in Critical Cleaning  
30 Glenn St. White Plains NY 10603 USA  
Tel 914.948.4040 • Fax 914.948.4088  
www.alconox.com • clean.ng@alconox.com