



Alconox, Inc.

The Leader in Critical-Cleaning

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 10/18/2017

CERTIFICATE OF ANALYSIS LUMINOX

Manufactured by: Alconox, Inc.

Sub lot No: M1K7 M1K7G, M1K7J, M1K7L, M1K7M

Sub lot numbers of Master lot No: M1H6

Manuf. Cat. no.: 1901, 1901-1, 1915, 1955

Date Of Manufacture: 10/2/2017

Expires end of: 10/19

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

<u>ANALYSIS</u>	<u>RESULT</u>	<u>ACCEPTED RANGE</u>
Appearance:	OK	Clear liquid
pH: (as is)	7.2	6.0 - 8.0
Surface Tension:(1% solution Dyne/cm)	33.3	32.0 - 38.0
Specific Gravity:	1.024	1.00 - 1.06
Foam (1% in DI water)	NIL	NIL

Malcolm McLaughlin, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.