

Alconox Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA Tel.914.948.4040 ◆ Fax.914.948.4088 www.alconox.com ◆ cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 3/30/2023

CERTIFICATE OF ANALYSIS TERGAJET

Manufactured by: Alconox Inc.

Sub lot No: T1C3 T1C3A, T1C3B, T1C3C, T1C3D

Sub lot numbers of Master lot No: T1A2

Manuf. Cat. no.: 2201, 2203, 2204, 2204-1, 2225, 2250

Date Of Manufacture: 3/28/2023

Expires end of: 03/25

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

| ANALYSIS | RESULT | ACCEPTED RANGE |
|---|---------------|-----------------------|
| Appearance: | OK | White powder |
| Solubility: (1% solution at 25 deg C, with agitation) | COMP | Complete |
| pH:(1% solution in DI water) | 11.3 | 11.0 - 12.5 |
| Surface Tension:(1% solution Dyne/cm) | 33.7 | 31.0 – 36.0 |
| Foam (1% in DI water) | NIL | NIL |

Laura Falla, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.

COA NUMBER: 8877